



UKRAINE
VOLYN OBLAST STATE ADMINISTRATION
HEALCARE DEPARTMANT

Stepana Bandery St., 5, Lutsk, 43025, tel./fax (0332) 241581, 243 559
e-mail: uoz@uoz.volyn.ua, Code 23254447

04.09.11 No. 2854/15/2-11

For No. _____ of _____

To all potential bidders

Amendments No.1
to the Bidding Documents for
ICB-G № 6.2.2.1 Procurement of medical equipment and consumables for
healthcare facilities of the secondary and tertiary levels (phase 1)

Dear Sirs and Madams!

According to para. 8 Section I "Instruction to Bidders" we would like to inform you that the following amendments (corrections) have been made to the Bidding Documents:

Lot No.1

The following amendments were made to technical requirements of sub para. 3.7 Detailed Requirements (technical specifications), para. 3 Technical Specifications, Section VII. Schedule of Requirements of Bidding Documents:

1) Item 5. **Set for daily blood pressure monitoring**

No.	Detailed requirements (old wording)	Detailed requirements (new wording)
1.12	Recording device size, not more than 130*70*40 mm	Recording device size, not more than 130* 75 *40 mm
1.14	Possibility to measure blood pressure in the range from 0 to 290 mmHg with a time interval of 15, 30, 60, 120, 240 min	Possibility to measure blood pressure in the range from 290 mmHg. with changeable a time interval
1.15	Recording device memory capacity, not less than 600 measurements	Recording device memory capacity, not less than 300 measurements

2) Item 6. **Bicycle ergometry kit**

No.	Detailed requirements (old wording)	Detailed requirements (new wording)
3.2	Dimensions, not more than 130x70 x40 mm	Dimensions, not more than 130x 90 x40 mm
3.13	Sensitivity, not worse than 5; 10; 20 mm/mV	Sensitivity, not worse than 2 ; 5; 10; 20; 25 mm/mV

Lot No.3

The following amendments were made to technical requirements of sub para. 3.7 Detailed Requirements (technical specifications), para. 3 Technical Specifications, Section VII. Schedule of Requirements of Bidding Documents (**English version**):

1) Item 1. **Hematology analyzer (type 1)**

No.	Detailed requirements (old wording)	Detailed requirements (new wording)
5	Blood sample volume of 10 mL	Blood sample volume up to 10 mL

Please confirm the receipt of this document "**Amendment No.1**" to e-mail volyn.wbproject@gmail.com.

Head



Vashchenyuk I.S

Performer Nesterenko
(0332) 777-207